

## OFFER FORM ACTED Libya

Date: التاريخ \_\_\_\_\_ (to be indicated by the bidder)

Tender N°: عدد المناقص: T/14DPE/H94/QIP/BENHAZI/MEDICAL/PRG/11062019/004

### To be Filled by Bidder (COMPULSORY)

#### Details of Bidding Company t/ تفاصيل الشركة المتقدمة

1. Company Name / اسم الشركة ( \_\_\_\_\_ )
2. Company Authorized Representative Name / اسم الشخص المفوض ( \_\_\_\_\_ )
3. Company Registration No / رقم السجل القانوني للشركة ( \_\_\_\_\_ )  
No/Country/ Ministry
4. Company Specialization / تخصص الشركة ( \_\_\_\_\_ )
5. Mailing Address / عنوان البريد ( \_\_\_\_\_ )  
Country/Governorate/City/St name/Shop-Office No
  - a. Contact Numbers / رقم الاتصال (Land Line: \_\_\_\_\_ / Mobile No: \_\_\_\_\_ )
  - b. E-mail Address / البريد الالكتروني ( \_\_\_\_\_ )

I undersigned \_\_\_\_\_, agree to provide ACTED, non-profit NGO, with items answering the following specifications, according to the general conditions and responsibilities that I engage myself to follow.

انا الموقع ادناه ----- اوافق على تزويد أكتد المنظمة الغير حكومية والغير ربحية بهذه البنود لتتلقى المواصفات التالية وفقا للشروط العامة والمسؤوليات التي وافقت عليها بنفسى.

**LOT 1:** MEDICAL EQUIPMENT AND INSTALLATION :

DDP: **BODRISA SCHOOL – Benghazi, Libya**, BENGHAZI, BENGHAZI, LIBYA // **AL LATHAMA HEALTH UNIT**, BENGHAZI, BENGHAZI, LIBYA // **AL JAEDI (JAEDA) CLINIC – EL LITHI**, BENGHAZI, BENGHAZI, LIBYA // **AL BU ATNI HEALTH UNIT**

**THE OFFER WILL BE SUBMITTED IN LIBYAN DINARS OR USD ALL COST INCLUDED (INCLUDING DELIVERY AND INSTALLATION) AND TAXES INCLUDED.**

Item No.	Item Description	DDP	Unit	Quantity	Unit Price LYD VAT Included	TOTAL Price LYD VAT Included	Unit Price USD VAT Included	TOTAL Price USD VAT Included
1- Medical Equipment and Installation								
1,1	ECG channels machine with screen, IEC class I, type CF, Power supply: AC:220V/110V, 50/60Hz, DC:12V rechargeable battery.	Benghazi Al Jadedda clinic - El Lithi - Benghazi	Pce	1				
1,2	ECHO 2D Mode and Doppler (PW-CW and color) machine (additional specs attached in <b>ANNEX 1</b> )	Benghazi Al Jadedda clinic - El Lithi - Benghazi	Pce	1				
1,3	Blood pressure measurer (LCD bar with backlight, Digital pressure read out on LCD, measure mode upper arm type manual inflation, 2M (1.5 V) alkaline battery, 2007 edition at least)	Buatni Clinic- Buatni - Benghazi;	Pce	4				
1,4	Blood pressure measurer (LCD bar with backlight, Digital pressure read out on LCD, measure mode upper arm type manual inflation, 2M (1.5 V) alkaline battery, 2007 edition at least)	Al Lathama health unit - Al Lathama - Benghazi	Pce	2				

1,5	Medical Bed, sleep surface of 36'' W x 80''L, according to hospital standards	Buatni Clinic- Buatni - Benghazi;	Pce	2					
1,6	Medical Bed, sleep surface of 36'' W x 80''L, according to hospital standards	Al Lathama health unit - Al Lathama - Benghazi;	Pce	2					
1,7	First Aid Kit (Kit contains essential first-aid supplies to treat most common injuries, including cuts, scrapes, swelling, sprains, strains and more). Thank you to attach a <b>BOM of the First aid Kit.</b>	Bodrisa school - Bodrisa – Benghazi	Pce	2					
<b>Grand Total (including all taxes, delivery and installation)</b>						<b>In LYD</b>		<b>In USD</b>	

Technical Specification of each one of those items should be provided attached with the bid.

**BIDDER'S COMMENTS/REMARKS:**

1. \_\_\_\_\_
2. \_\_\_\_\_

**DELIVERY CONDITIONS GUARANTEED BY THE BIDDER :**

..... Days for 100% of the items

**BIDDER'S TERMS AND CONDITIONS:**

1. Valid of the offer: \_\_\_\_\_ (recommended: 3 months or more)
2. Terms of delivery: \_\_\_\_\_ (DDP recommended)



T/14DPE/H94/QIP/BENHAZI/MEDICAL/PRG/11062019/004

- 3. Terms of payment: \_\_\_\_\_
- 4. Payments deadline: \_\_\_\_\_ days after invoice validated

Note: The offer will be submitted in English and in Libyan Dinars or USD all cost included (including delivery and installation) and taxes included.

- a. Bidders shall comply with all valid Government legal Documentation and adhere to Regulations to operate in Libya such as being regular tax payer to offer such goods and services as specified in the tender

- 1. An entity established in Libya is considered tax resident in Libya.
  - a. Any Income generated in Libya from work performed therein should be subject to income tax in Libya.
  - b. The supplier will have the obligation to provide Tax information, including handling the preparation, filing, and stamping of federal, state, or national body on the invoice of the performed service/work.

Name of Bidder's Authorized Representative: \_\_\_\_\_

Authorized signature and stamp: \_\_\_\_\_

Date: \_\_\_\_\_



## ANNEX 1 –

**ECHOCARDIOGRAPHY****TECHNICAL SPECIFICATIONS**

1. Should be a stand alone system integrated on a light weight mobile cart.
2. The system should be a color Doppler Echocardiography all digital beam former system to study the anatomical abnormalities and blood flow in the heart and associated vessels. Should be a stand alone system integrated on a light weight mobile cart.
3. Should be a latest generation Electronic Phased array Color Doppler system with minimum 512 Electronic independent channels.
4. Should have 256 gray shades for sharp contrast resolutions.
5. Should be supplied with adult and pediatric cardiac and vascular probes of wide band transducers without frequency selection for higher sensitivity of response over a broad frequency range of operation.
6. Should have 2D, M-mode, Anatomical M-mode, Color M-mode, PW and CW doppler, Steerable CW doppler.
7. The system should have a very high dynamic range of at least 200dB to pick up subtle echoes.
8. Should have three active ports.
9. Should have 2-4 Mhz broadband phased array sector probe for adult cardiac imaging.
10. Should have 3-8 Mhz broadband phased array sector probe for paediatric and neonatal cardiac imaging.
11. Should have 3-12 Mhz broadband Linear Array probe for vascular imaging.
12. Should have multi frequency convex array probe 3-10 MHz for paediatric imaging
13. Pencil probe (optional)
14. Should have advanced tissue Harmonic Imaging.
15. Should have color flow imaging.
16. Should have color Tissue Doppler Imaging.
17. Should have gain control in Axial Plane.
18. Should have triple imaging possibility on the system.
19. Should have PW/CW Doppler facility in all imaging phased array sector probes.
20. Should have 15" or more high resolution TFT monitor with tilt and swivel facility and should be able to view in all angles and all light conditions.
21. Should have greater than 5000 images in the system hard disk drive
22. Should have in built CD/DVD writer.
23. Should have patient reporting page with embedded images.



*T/14DPE/H94/QIP/BENHAZI/MEDICAL/PRG/11062019/004*

24. Should have full functional measurement facility and calculation should be possible.
25. Should be supplied with thermal printer and 6 packs of thermal paper and the unit should have option to connect external printer.
26. Unit should function with 200-240Vac, 50/60 Hz input power supply.
27. Unit should be supplied with suitable UPS with a minimum 30 minutes back-up time.
28. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.